



Ref. Certif. No.

**DK-45833-UL**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**Product  
Produit

Switch Mode Power Supply

Name and address of the applicant  
Nom et adresse du demandeurTDK-LAMBDA UK LTD  
KINGSLEY AVE  
ILFRACOMBE, EX34 8ES United KingdomName and address of the manufacturer  
Nom et adresse du fabricantTDK-LAMBDA UK LTD  
KINGSLEY AVE  
ILFRACOMBE, EX34 8ES United KingdomName and address of the factory  
Nom et adresse de l'usinePanyu Trio Microtronics Co Ltd  
SHIJI INDUSTRIAL ESTATE  
DONGYONG  
NANSHA  
GUANGZHOU, 511453 GUANGDONG ChinaNote: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>eme</sup> page Additional Information on page 2  
94.5 Vac to 240 Vac, (85-264Vac max. tolerance),  
16A, 47-63 HzRatings and principal characteristics  
Valeurs nominales et caractéristiques principales

TDK-Lambda

Trademark (if any)  
Marque de fabrique (si elle existe)**TDK-Lambda**Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

SMT

Model / Type Ref.  
Ref. De typeSeries: Alpha 800, Alpha 800W. Models: CA800 (followed by  
various letters and numbers as defined in the model differences)  
Series: Alpha 1000, Alpha 1000W., See Page 2Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>eme</sup> pageAdditionally evaluated to EN 60601-1:2006 / A1:2013; National  
Differences specified in the CB Test Report Additional Information on page 2A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part  
of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

E349607-D8-CB-1 issued on 2015-05-27

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2015-05-27

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-45833-UL**

**Model Details:**

Models: CA1000 (followed by various letters and numbers as defined in the model differences), CA1250 12C\_MF\_PP 12F\_PP 12F\_PP 12F\_PP, CA1250 12C\_MF 12FF 12FF 12FF

**Factories:**

TDK-LAMBDA UK LTD  
KINGSLEY AVE  
ILFRACOMBE, EX34 8ES United Kingdom

**Additional Information:**

No Risk Management evaluation has been carried out on this product, therefore Essential Performance has not been specified, as this depends on the End Equipment

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/nbcnames](http://www.ul.com/nbcnames)

Date: 2015-05-27

Signature:   
Jan-Erik Storgaard