



Ref. Certif. No.

DK-107035-M2-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product: Switching Power Supply

Name and address of the applicant: TDK-LAMBDA CORP, NAGAOKA TECHNICAL CENTER, R&D DIV, 2704-1 SETTAYA-MACHI NAGAOKA-SHI, NIIGATA, 940-1195 Japan

Name and address of the manufacturer: TDK-LAMBDA CORP, NAGAOKA TECHNICAL CENTER, R&D DIV, 2704-1 SETTAYA-MACHI NAGAOKA-SHI, NIIGATA, 940-1195 Japan

Name and address of the factory: TDK-LAMBDA CORP, 2704-1 SETTAYA-MACHI NAGAOKA-SHI, NIIGATA-KEN, 940-1195 Japan

Note: When more than one factory, please report on page 2

Ratings and principal characteristics: Input: 100 - 240Vac, 50-60Hz, 6.2A

Additional Information on page 2

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.: CME350P-1000-abcdef, CUS350MP-1000-abcdef

Additional Information on page 2

Additional information (if necessary may also be reported on page 2):

Additionally evaluated to:
 EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014
 National Differences specified in the CB Test Report.
 The report was revised to include technical modifications.
 The risk management requirements of the standard were not addressed

Additional Information on page 2

A sample of the product was tested and found to be in conformity with: IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate: E309264-D1050-1/A2/C0-CB issued on 2022-03-31

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-04-18
 Original Issue Date: 2020-12-11

Signature: Jan-Erik Storgaard



Ref. Certif. No.

DK-107035-M2-UL

Factory(ies):

TDK-LAMBDA MALAYSIA SDN BHD
LOT 2 & 3, BATU 9 3/4
KAWASAN PERINDUSTRIAN
BANDAR BARU JAYA GADING KUANTAN, PAHANG, 26070
Malaysia

TDK-LAMBDA MALAYSIA SDN BHD
PLO33 KAWASAN PERINDUSTRIAN SENAI SENAI, JOHOR, 81400
Malaysia

TDK-Lambda (China) Electronics Co Ltd
No.95,Zhujiang Rd, Xinwu District Wuxi, JiangSu, 214028
China

Additional Model Detail(s):

CME350P-1000-abcdef, CUS350MP-1000-abcdef,
where a is 24, 30, 36 or 48 and "bcdef" is set of optional suffix (b is /, c is T, d is A or L, e is CO2, f is SF, and "b", "c", "d", "e"
and "f" maybe blank.

Additional Ratings:

Output:

<CME350P-1000-24 / CUS350MP-1000-24> Nominal: 24Vdc (23.5-26.9Vdc), 14.6A (Peak: 24Vdc, 41.7A)
<CME350P-1000-30 / CUS350MP-1000-30> Nominal: 30Vdc (26.5-30.5Vdc), 11.65A (Peak: 30Vdc, 33.3A)
<CME350P-1000-36 / CUS350MP-1000-36> Nominal: 36Vdc (35.5-42.5Vdc), 9.7A (Peak: 36Vdc, 27.7A)
<CME350P-1000-48 / CUS350MP-1000-48 > Nominal: 48Vdc (44.5-48.5Vdc), 7.3A (Peak: 48Vdc, 20.9A)

Summary of Modifications:

Deleted one factory.; Added alternate components.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-04-18
Original Issue Date: 2020-12-11

Signature: 
Jan-Erik Storgaard