



Ref. Certif. No.

US-43136-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Power Supply
Name and address of the applicant	TDK-LAMBDA CORP NAGAOKA TECHNICAL CENTER R&D DIV 2704-1 SETTAYA-MACHI NAGAOKA-SHI, Niigata, 940-1195 Japan
Name and address of the manufacturer	TDK-LAMBDA CORP NAGAOKA TECHNICAL CENTER R&D DIV 2704-1 SETTAYA-MACHI NAGAOKA-SHI, Niigata, 940-1195 Japan
Name and address of the factory	TDK-LAMBDA MALAYSIA SDN BHD PLO33 KAWASAN PERINDUSTRIAN SENAI SENAI, Johor 81400 Malaysia <input checked="" type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input: 100-240 VAC, 7.0 A, 50-60 Hz, Output: - GXE600-24: 24 VDC, 25 A, - GXE600-48: 48 VDC, 12.5 A
Trademark / Brand (if any)	TDK-Lambda
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	GXE600-24, GXE600-48 <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021 The risk management requirements of the standard were not addressed National Differences: EU Group Differences, CA, US <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E309264-D1074-1/A0/C0-CB issued on 2023-12-06

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-12-12

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-43136-UL

Factory(ies):

TDK-LAMBDA MALAYSIA SDN BHD
LOT 2 & 3, BATU 9 3/4
KAWASAN PERINDUSTRIAN
BANDAR BARU JAYA GADING KUANTAN, Pahang 26070
Malaysia

TDK-LAMBDA CORP
2704-1 SETTAYA-MACHI NAGAOKA-SHI, Niigata, 940-1195
Japan

TDK-Lambda (China) Electronics Co Ltd
No.95, Zhujiang Rd, Xinwu District Wuxi, Jiangsu Sheng, 214028
China

Additional Model Detail(s):

GXE600-24, GXE600-48, may be followed by suffix "abcdefg" (a is /, b is HD, c is A, d is FG, e is SF, f is NS, g is NHV for GXE600-48, and "a", "b", "c", "d", "e", "f", "g" may be blank) (for the details, see General Product Information)

Additional information (if necessary)



- UL Solutions (US), 333 Pflingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-12-12

Signature:

Jolanta M. Wroblewska