



Ref. Certif. No.

DK-5219-A2-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Component Power Supply

Name and address of the applicant

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM

Name and address of the manufacturer

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM

Name and address of the factory

Note: When more than one factory, please report on page 2

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM

Additional Information on page 2

Ratings and principal characteristics

100-240Vac nominal (90-264V max. tolerance), 47-63Hz, 5.5A
(See Test Report for details of model ratings)

Trademark (if any)

TDK-Lambda



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

NV350, NV-350, NF350 or NV3 Range, See Page 2

Additional information (if necessary may also be reported on page 2)

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.2), IEC 60601-1(ed.2);am1, IEC 60601-1(ed.2);am2

As shown in the Test Report Ref. No. which forms part of this Certificate

E349607-A6-CB-1 issued on 2017-05-22

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-05-22

Original Issue Date: 2011-12-01

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-5219-A2-UL

Model Details:

NV-350, NV350

(may be prefixed by NS - # / or - where # may be up to any four letters and may be followed by - \$; where \$ maybe any number between 000 to 999, indicating non-safety related model differences.)

Factories:

PANYU TRIO MICROTRONIC CO LTD
SHIJI INDUSTRIAL ESTATE DONGYONG NANSHA GUANGZHOU GUANGDONG
CHINA

Additional Information:

National Differences specified in the CB Test Report.

The original report was modified to include the following changes/additions:

- Updated critical components;
- Revise models.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/names

Date: 2017-05-22

Original Issue Date: 2011-12-01

Signature:

Jan-Erik Storgaard