

Certificate



Certificate no.

CU 72226004 01

License Holder:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Manufacturing Plant:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Test report no.: USA-SOS CN22ADLF 001

Client Reference: GORINSHTEIN BORIS

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: Programmable AC/DC power supply

License Fee - Units

Model Designation: 1a) Gxxxx-yyy-z-v/uuuuuu/w or
Gxxxx-yyy-z-v-uuuuuu/w (Ordinary unit) 8
1b) GBxxxx-yyy-z-v/uuuuuu/w or GBxxxx-yyy-
GBxxxx-yyy-z-v-uuuuuu/w (Blank units)
2a) G1000-5-z-v/uuuuuu/w or
G1000-5-z-v-uuuuuu/w (Ordinary unit)
2b) GB1000-5-z-v/uuuuuu/w or
GB1000-5-z-v-uuuuuu/w (Blank units)
3a) G1500-3.4-z-v/uuuuuu/w or
G1500-3.4-z-v-uuuuuu/w (Ordinary unit)
3b) GB1500-3.4-z-v/uuuuuu/w or Contd.

8

Appendix: 1

Licensed Test mark:



Date of Issue

(day/mo/yr)
20/05/2022

Certificate



Certificate no.

CU 72226004 02

License Holder:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Manufacturing Plant:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Test report no.: USA-SOS CN22ADLF 001

Client Reference: GORINSHTEIN BORIS

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: Programmable AC/DC power supply

License Fee - Units

Contd.

Model Designation: GB1500-3.4-z-v-uuuuuu/w (Blank units)
xxxx=020 - 1500;yyy=005 - 375;z=GPIB (IEEE),MDBS,ECAT,IS420
IS010 or Blank;v=3P200 or 3P480;uuuuuu=Various letters and /
or numbers or Blank;w=CO or Blank

Rating for input :Option 1: AC 190-240 V; 47-63 Hz; 3W+PE.
1a), 1b): Rated input current: 27.8 A max.
2a), 2b): Rated input current: 18.5 A max.
3a), 3b): Rated input current: 18.5 A max.
Option 2: AC 380-480 V; 47-63 Hz; 3W+PE.
1a), 1b): Rated input current: 13.8 A max.
Contd.

Licensed Test mark:



Date of Issue

(day/mo/yr)
20/05/2022

Certificate



Certificate no.

CU 72226004 03

License Holder:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Manufacturing Plant:

TDK-Lambda Ltd.
56 Haharoshet St., P.O.B. 500
Karmiel Industrial Zone
2161401 Karmiel
Israel

Test report no.: USA-SOS CN22ADLF 001

Client Reference: GORINSHTEIN BORIS

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: Programmable AC/DC power supply

License Fee - Units

Contd.

Rating for input :2a), 2b): Rated input current: 9.2 A max.
3a), 3b): Rated input current: 9.2 A max.
Ratings for output:1a), 1b): DC 0-20V/375A to DC 0-1500V/5A,
7500 Watt max.
2a), 2b): DC 0-1000V/5A, 5000 Watt.
3a), 3b): DC 0-1500V/3.4A, 5100 Watt.

Class of equipment:Class I

Licensed Test mark:



Date of Issue

(day/mo/yr)
20/05/2022