



Ref. Certif. No.

DE 2-032869

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product	Programmable Power Supply
Name and address of the applicant	TDK-Lambda Ltd. Karmiel Industrial Zone 56 Haharoshet St., P.O.B. 500, 2161401 Karmiel, Israel
Name and address of the manufacturer	TDK-Lambda Ltd. Karmiel Industrial Zone 56 Haharoshet St., P.O.B. 500, 2161401 Karmiel, Israel
Name and address of the factory	TDK-Lambda Ltd. Karmiel Industrial Zone 56 Haharoshet St., P.O.B. 500, 2161401 Karmiel, Israel
Ratings and principal characteristics	1),2),3),6),7) Option 1: AC 190-240V; 3W + PE, 50/60Hz; 1),2),3),6),7) Option 2: AC 380-415V; 3W + PE, 50/60Hz; 1),2),3),6),7) Option 3: AC 380-480V; 3W + PE, 50/60Hz; Class I (see page 2 and test report for further ratings)
Trademark (if any)	TDK-Lambda
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	1) GENESYS+5000W series; 2) GSP/GBSP 10kW series; 3) GSP/GBSP/ GSSP 15KW series; 4) GENESYS+ 1700W series; 5) GENESYS+ GH1500W series; 6) GENESYS+ 2700W series; 7) GENESYS+ 3400W series; 8) GENESYS+ 1000W series; 9) GENESYS+ GH1000W series (see page 3 for details)
Additional information (if necessary may also be reported on page 2)	Replaces CB Certificate US-TUVR-12197 due to change of NCB.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31781623 303

This CB Test Certificate is issued by the National Certification Body

TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2021-10-25

Signature:

Dipl.-Ing. M. Glagla



Date: \_\_\_\_\_ Signature: *Martin Hagg*

10/061a CB 01/21 rk



Date: \_\_\_\_\_ Signature: *Martin Hagg*

10/061a CB 01/21 rk